

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 103

Burettes, 10 ml, str. stopc., PTFE-plug, AS, conf. cert.

Batch: 26-24




Serial:

Measuring Nr. 73425

Batch size:

91

Volume:	10,0000 ml	6,0000 ml	2,0000 ml
Tolerance:	$\pm 0,0200$ ml	$\pm 0,0200$ ml	$\pm 0,0200$ ml
Average value:	10,0014 ml	5,9968 ml	2,0019 ml
Deviation:	0,0049 ml	0,0062 ml	0,0024 ml
Precision (CV):	0,0491 %	0,1041 %	0,1210 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

25.06.2025

Balance:

Sartorius -
LE244S-OCE

DE-M