

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 025

Graduated cylinder, 250:2 ml, A, conf. cert.




Batch: 17-25

Serial:

Measuring Nr. 72814

Batch size: 500

Volume:	50,0000 ml	150,0000 ml	250,0000 ml
Tolerance:	$\pm 1,0000$ ml	$\pm 1,0000$ ml	$\pm 1,0000$ ml
Average value:	49,8833 ml	150,0865 ml	250,1576 ml
Deviation:	0,1780 ml	0,1564 ml	0,1530 ml
Precision (CV):	0,3561 %	0,1043 %	0,0612 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

24.04.2025

Balance:

Sartorius - LA
1200 S

DE-M