

# Certificate of performance



**Article:** 3 662 000 BL

**Vol. flasks, 5 ml, NS 7, blue grad., A, conf. cert.**

**Batch:** 24-25

**Serial:**

-

**Measuring Nr.** 73240

**Batch size:**

500


**Volume:** 5,0000 ml

**Tolerance:**  $\pm 0,0250$  ml

**Average value:** 4,9976 ml

**Deviation:** 0,0039 ml

**Precision (CV):** 0,0775 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

16.06.2025

**Balance:**

Sartorius -  
LE244S-OCE

