

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 002 USP

**Volumetric flasks, 20 ml, NS 10, USP, A, conf. cert.**

**Batch:** 15-25


**Serial:**

**Measuring Nr.** 72785

**Batch size:**

502

|                        |                 |
|------------------------|-----------------|
| <b>Volume:</b>         | 20,0000 ml      |
| <b>Tolerance:</b>      | $\pm 0,0300$ ml |
| <b>Average value:</b>  | 20,0039 ml      |
| <b>Deviation:</b>      | 0,0031 ml       |
| <b>Precision (CV):</b> | 0,0154 %        |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

16.04.2025

**Balance:**

Sartorius -  
LE244S-OCE

DE-M