

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 005

Vol. flasks, 50 ml, NS 12, A, conf. cert.


Batch: 17-25

Serial:

Measuring Nr. 72980

Batch size: 2983

Volume:	50,0000 ml
Tolerance:	$\pm 0,0600$ ml
Average value:	50,0012 ml
Deviation:	0,0052 ml
Precision (CV):	0,0105 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

15.05.2025

Balance:

Sartorius -
LE244S-OCE

DE-M