

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 0075

**Volumetric flask, 75 ml, NS 12, A, conf. cert.**

**Batch:** 29-25


**Serial:**

**Measuring Nr.** 73538

**Batch size:**

146

|                        |                   |
|------------------------|-------------------|
| <b>Volume:</b>         | <b>75,0000 ml</b> |
| <b>Tolerance:</b>      | $\pm 0,0600$ ml   |
| <b>Average value:</b>  | 74,9953 ml        |
| <b>Deviation:</b>      | 0,0042 ml         |
| <b>Precision (CV):</b> | 0,0055 %          |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

08.07.2025

**Balance:**

Sartorius -  
LE244S-OCE

DE-M