

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 090 BL

Vol. flask., 900 ml, NS 24, blue grad., A, conf. cert.

Batch: 10-25


Serial:

Measuring Nr. 72300

Batch size:

100

Volume:	900,0000 ml
Tolerance:	$\pm 0,4000$ ml
Average value:	900,1282 ml
Deviation:	0,0308 ml
Precision (CV):	0,0034 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

13.03.2025

Balance:

Sartorius - LA
1200 S

DE-M