

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 200 BL

Vol. flask, 2000 ml, NS 29, blue grad., A, conf. cert.

Batch: 12-25

Serial:

Measuring Nr. 72361

Batch size:

279


Volume: 2000,0000 ml

Tolerance: $\pm 0,6000$ ml

Average value: 2000,2582 ml

Deviation: 0,0604 ml

Precision (CV): 0,0030 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

25.03.2025

Balance:

Sartorius -
MSE14202S-
OCE-D0

DE-M