

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 500

Volumetric flask, 5000 ml, NS 34, A, conf. cert.

Batch: 10-25

Serial:

Measuring Nr. 72458

Batch size:

60


Volume: 5000,0000 ml

Tolerance: $\pm 1,2000$ ml

Average value: 5000,3833 ml

Deviation: 0,2066 ml

Precision (CV): 0,0041 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

31.03.2025

Balance:

Sartorius -
MSE14202S-
OCE-D0

DE-M