# Certificate of performance

witeg GERMANY
ISO 9001 CERTIFIED

| | |
|---|---|
| **Article:** | **3 669 001 BL** |
| **Vol. flask, 2 ml, NS 7, trapez., blue grad., A, conf. cert.** | |

| | | | |
|---|---|---|---|
| **Batch:** | 38-26 | **Serial:** | |
| **Measuring Nr.** | 79960 | **Batch size:** | 278 |

| | |
|---|---|
| **Volume:** | **2,0000 ml** |
| **Tolerance:** | ± 0,0250 ml |
| **Average value:** | 2,0014 ml |
| **Deviation:** | 0,0026 ml |
| **Precision (CV):** | 0,1276 % |




These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**
Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

| | |
|---|---|
| Operator: | 7 |
| Date: | 16.09.2026 |
| Balance: | Sartorius - LE244S-0CE |

DE-M