

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 002

Vol. flask, 20 ml, NS 10, amber glass, A, conf. cert.

Batch: 02-25

Serial:

Measuring Nr. 72082

Batch size: 499


Volume: 20,0000 ml

Tolerance: $\pm 0,0400$ ml

Average value: 19,9993 ml

Deviation: 0,0041 ml

Precision (CV): 0,0206 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

03.03.2025

Balance:

Sartorius -
LE244S-OCE

DE-M