

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 002

Vol. flask, 20 ml, NS 10, amber glass, A, conf. cert.


Batch: 27-26

Serial:

Measuring Nr. 79285

Batch size: 985

Volume:	20,0000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	19,9981 ml
Deviation:	0,0055 ml
Precision (CV):	0,0277 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

30.07.2026

Balance:

Sartorius -
LE244S-OCE

DE-M