

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 011 USP

Vol. flask, 100 ml, amber glass, NS 14, A, conf. cert.

Batch: 28-26


Serial:

Measuring Nr. 79517

Batch size:

279

Volume:	100,0000 ml
Tolerance:	$\pm 0,0800$ ml
Average value:	100,0307 ml
Deviation:	0,0218 ml
Precision (CV):	0,0218 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

10.08.2026

Balance:

Sartorius -
MSE224S-0CE
-DU

DE-M