

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 050

Volumetric pipettes, 5 ml, AS, conf. cert.

Batch: 05-25


Serial:

Measuring Nr. 72332

Batch size:

2391

Volume:	5,0000 ml
Tolerance:	$\pm 0,0150$ ml
Average value:	5,0029 ml
Deviation:	0,0033 ml
Precision (CV):	0,0668 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

18.03.2025

Balance:

Sartorius -
LE244S-OCE

DE-M