

Graduated pipettes

USP

- color-code to avoid confusion
- made of soda-lime glass
- zero point at the upper end of the scale
- main point ring graduation
- tolerance according to ASTM E 1293
- produced according to DIN EN ISO 835
- DE-M marked according to the German calibration law
- calibrated to deliver (EX), 5 sec waiting time
- inclusive batch certificate



Capacity ml	Division ml	Tolerance USP ml	Color-Code	Order number	
				blue graduated	brown graduated
1	0.01	±0.01	yellow	3 510 101 BL USP	3 510 101 USP
2	0.02	±0.01	black	3 510 202 BL USP	3 510 202 USP
5	0.05	±0.02	red	3 510 505 BL USP	3 510 505 USP
5	0.1	±0.02	blue	3 510 510 BL USP	3 510 510 USP
10	0.1	±0.03	orange	3 510 710 BL USP	3 510 710 USP

Class AS - enzymatic analysis

- color-coded to avoid confusion
- made of soda-lime glass
- zero point at the upper end of the scale
- main point ring graduation
- tolerance according to DIN EN ISO 835 standard
- DE-M marked according to the German calibration law
- calibrated to deliver (EX)
- according to DIN 12699
- inclusive batch certificate



Capacity ml	Division ml	Tolerance ml	Color-Code	Delivery time seconds	Order number	
					blue graduated	brown graduated
0.1	0.001	0.003	green x3	2 - 8	3 517 011 BL	3 517 011
0.2	0.001	0.003	blue x3	2 - 8	3 517 021 BL	3 517 021
0.5	0.01	0.005	yellow x3	2 - 8	3 517 051 BL	3 517 051
1	0.01	0.006	yellow x2	2 - 8	3 517 101 BL	3 517 101
2	0.02	0.010	black x2	2 - 8	3 517 202 BL	3 517 202
5	0.05	0.030	red x2	5 - 11	3 517 505 BL	3 517 505
10	0.1	0.050	orange x2	5 - 11	3 517 710 BL	3 517 710

Pipette stand, universal use, detachable

made of PP, for maximum of 94 pipettes, specially designed to prevent any damage to the pipette tip. Residual liquids are retained by reinforced rim.

Order number

7 081 001

