

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 010 102

Burettes, 10 ml, straight stopc., AS, conf. cert.




Batch: 11-22

Serial:

Measuring Nr. 56374

Batch size: 10

Volume:	10,0000 ml	6,0000 ml	2,0000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	9,9995 ml	5,9963 ml	1,9986 ml
Deviation:	0,0064 ml	0,0072 ml	0,0029 ml
Precision (CV):	0,0637 %	0,1194 %	0,1454 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

31.05.2022

Balance:

Sartorius -
LE244S-OCE

DE-M