

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 010 103 B

**Burettes, 10 ml, str. stopc., amber grad., AS, conf. cert.**




**Batch:** 38-15

**Serial:**

**Measuring Nr.** 46703

**Batch size:** 150

Volume:	10,000 ml	6,000 ml	2,000 ml
<b>Tolerance:</b>	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
<b>Average value:</b>	10,0070 ml	6,0029 ml	1,9999 ml
<b>Deviation:</b>	0,0034 ml	0,0041 ml	0,0028 ml
<b>Precision (CV):</b>	0,0343 %	0,0675 %	0,1411 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	12.05.2021
<b>Balance:</b>	Sartorius - LE244S-OCE

