

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 010 103

**Burettes, 10 ml, straight stopc., PTFE, AS, conf. cert.**




**Batch:** 05-20

**Serial:**

**Measuring Nr.** 36187

**Batch size:** 50

Volume:	10,0000 ml	6,0000 ml	2,0000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	10,0037 ml	5,9995 ml	2,0043 ml
Deviation:	0,0035 ml	0,0036 ml	0,0037 ml
Precision (CV):	0,0349 %	0,0606 %	0,1863 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	9
Date:	19.02.2020
Balance:	Sartorius - LE244S-OCE

DE-M