

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 010 503

Burettes, 50 ml, str. stopcock, PTFE, AS, conf. cert.




Batch: 43-21

Serial:

Measuring Nr. 55423

Batch size: 120

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	49,9999 ml	29,9756 ml	9,9987 ml
Deviation:	0,0141 ml	0,0093 ml	0,0129 ml
Precision (CV):	0,0283 %	0,0310 %	0,1292 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	13.05.2022
Balance:	Sartorius - LE244S-OCE

