

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 052

Burettes, 5 ml, straight stopc., glass, AS, conf. cert.




Batch: 25-21

Serial:

Measuring Nr. 59621

Batch size: 10

Volume:	5,0000 ml	3,0000 ml	1,0000 ml
Tolerance:	± 0,0100 ml	± 0,0100 ml	± 0,0100 ml
Average value:	4,9972 ml	2,9956 ml	0,9997 ml
Deviation:	0,0024 ml	0,0008 ml	0,0012 ml
Precision (CV):	0,0475 %	0,0256 %	0,1172 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	77
Date:	06.10.2022
Balance:	Sartorius - LE244S-OCE

