

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 053

Burettes, 5 ml, str. stopc., PTFE plug, AS, conf. cert.

Batch: 24-16




Serial:

Measuring Nr. 15909

Batch size:

1

Volume:	5,0000 ml	3,0000 ml	1,0000 ml
Tolerance:	± 0,0100 ml	± 0,0100 ml	± 0,0100 ml
Measured value:	5,0058 ml	2,9959 ml	1,0022 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

9

Date:

23.11.2017

Balance:

Sartorius -
LE244S-OCE

DE-M