

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 053

Burettes, 5 ml, str. stopc., PTFE- plug, AS, conf. cert.




Batch: 07-26

Serial:

Measuring Nr. 77798

Batch size: 20

Volume:	5,0000 ml	3,0000 ml	1,0000 ml
Tolerance:	$\pm 0,0100$ ml	$\pm 0,0100$ ml	$\pm 0,0100$ ml
Average value:	5,0034 ml	3,0008 ml	1,0024 ml
Deviation:	0,0016 ml	0,0025 ml	0,0022 ml
Precision (CV):	0,0321 %	0,0840 %	0,2184 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

27.04.2026

Balance:

Sartorius -
LE244S-OCE

DE-M