

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 252 TR

Burette, 25 ml, with funnel, AS, conf. cert.




Batch: 38-15

Serial:

Measuring Nr. 14839

Batch size: 1

Volume:	25,0000 ml	15,0000 ml	5,0000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Measured value:	25,0000 ml	15,0075 ml	5,0064 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	11.10.2017
Balance:	Sartorius - LE244S-OCE

