

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 253

Burettes, 25 ml, str. stopc., PTFE plug, AS, conf. cert.




Batch: 41-17

Serial:

Measuring Nr. 15619

Batch size: 100

Volume:	25,000 ml	15,000 ml	5,000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	25,0004 ml	14,9932 ml	4,9962 ml
Deviation:	0,0095 ml	0,0057 ml	0,0056 ml
Precision (CV):	0,0380 %	0,0379 %	0,1119 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	09.11.2017
Balance:	Sartorius - LE244S-OCE

