

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 012 253

**Burettes, 25 ml, str. stopc., PTFE plug, AS, conf. cert.**




**Batch:** 11-23

**Serial:**

**Measuring Nr.** 63130

**Batch size:** 50

Volume:	25,0000 ml	15,0000 ml	5,0000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	25,0095 ml	15,0099 ml	5,0152 ml
Deviation:	0,0028 ml	0,0048 ml	0,0054 ml
Precision (CV):	0,0113 %	0,0323 %	0,1085 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	7
Date:	22.05.2023
Balance:	Sartorius - LE244S-OCE

DE-M