

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 503

Burettes, 50 ml, str. stopc., PTFE plug, AS, conf. cert.




Batch: 49-16

Serial:

Measuring Nr. 16630

Batch size: 24

Volume:	50,000 ml	30,000 ml	10,000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	49,9995 ml	29,9949 ml	9,9978 ml
Deviation:	0,0080 ml	0,0116 ml	0,0134 ml
Precision (CV):	0,0159 %	0,0387 %	0,1343 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	20.12.2017
Balance:	Sartorius - LE244S-OCE

