

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 012 503

**Burettes, 50 ml, str. stopc., PTFE plug, AS, conf. cert.**




**Batch:** 29-20

**Serial:**

**Measuring Nr.** 39442

**Batch size:** 70

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	50,0020 ml	29,9962 ml	10,0060 ml
Deviation:	0,0174 ml	0,0190 ml	0,0095 ml
Precision (CV):	0,0347 %	0,0634 %	0,0951 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	9
Date:	06.08.2020
Balance:	Sartorius - LE244S-OCE

