

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 012 503

Burettes, 50 ml, str. stopc., PTFE plug, AS, conf. cert.




Batch: 11-23

Serial:

Measuring Nr. 62748

Batch size: 198

Volume:	50,000 ml	30,000 ml	10,000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	49,9986 ml	30,0147 ml	10,0246 ml
Deviation:	0,0158 ml	0,0114 ml	0,0101 ml
Precision (CV):	0,0317 %	0,0380 %	0,1014 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	26.04.2023
Balance:	Sartorius - LE244S-OCE

