

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 014 103

Burettes, 10 ml, str. stopc., amber glass, AS, conf. cert.




Batch: 29-21

Serial:

Measuring Nr. 49643

Batch size: 20

Volume:	10,000 ml	6,000 ml	2,000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	9,9899 ml	5,9893 ml	1,9978 ml
Deviation:	0,0041 ml	0,0042 ml	0,0043 ml
Precision (CV):	0,0414 %	0,0694 %	0,2170 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 08.10.2021

Balance: Sartorius - LE244S-OCE

