

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 014 503

Burettes, 50 ml, str. stopc., amber glass, AS, conf. cert.




Batch: 38-17

Serial:

Measuring Nr. 15014

Batch size: 50

Volume:	50,000 ml	30,000 ml	10,000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	49,9840 ml	29,9780 ml	10,0074 ml
Deviation:	0,0096 ml	0,0018 ml	0,0072 ml
Precision (CV):	0,0192 %	0,0061 %	0,0717 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	17.10.2017
Balance:	Sartorius - LE244S-OCE

