

Certificate of performance



Article: 3 014 503

Burettes, 50 ml, str. stopc., amber glass, AS, conf. cert.




Batch: 43-21

Serial:

Measuring Nr.: 52131

Batch size: 80

Volume:	50,000 ml	30,000 ml	10,000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	50,0130 ml	30,0057 ml	10,0039 ml
Deviation:	0,0109 ml	0,0193 ml	0,0195 ml
Precision (CV):	0,0219 %	0,0643 %	0,1953 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	14.02.2022
Balance:	Sartorius - LE244S-OCE

