

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 014 503

Burettes, 50 ml, str. stopc., amber glass, AS, conf. cert.




Batch: 13-25

Serial:

Measuring Nr. 75260

Batch size: 30

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	50,0158 ml	29,9808 ml	9,9953 ml
Deviation:	0,0106 ml	0,0077 ml	0,0066 ml
Precision (CV):	0,0213 %	0,0257 %	0,0659 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	13.11.2025
Balance:	Sartorius - MSE224S-0CE -DU

