

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 022 107

Burettes, 10 ml, str. needle valve, AS, conf. cert.




Batch: 43-20

Serial:

Measuring Nr.: 45478

Batch size: 4

Volume:	10,000 ml	6,000 ml	2,000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	10,0050 ml	6,0015 ml	2,0029 ml
Deviation:	0,0014 ml	0,0020 ml	0,0041 ml
Precision (CV):	0,0144 %	0,0335 %	0,2050 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	07.04.2021
Balance:	Sartorius - MSE224S-0CE -DU

