

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 122 105

Pellet-burettes, 10 ml, lat. needle valve, AS, conf. cert.




Batch: 13-18

Serial:

Measuring Nr.: 22100

Batch size: 9

Volume:	10,000 ml	6,0000 ml	2,0000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	10,0009 ml	5,9934 ml	2,0025 ml
Deviation:	0,0029 ml	0,0045 ml	0,0032 ml
Precision (CV):	0,0293 %	0,0755 %	0,1624 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	09.07.2018
Balance:	Sartorius - LE244S-OCE

