

# Certificate of performance



**Article:** 3 122 107

**Pellet-burettes, 10 ml, lat. needle valve, AS, conf. cert.**




**Batch:** 02-24

**Serial:**

**Measuring Nr.:** 66153

**Batch size:** 1

Volume:	10,000 ml	6,000 ml	2,000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Measured value:	9,9830 ml	6,0061 ml	1,9989 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	7
Date:	27.02.2024
Balance:	Sartorius - LE244S-OCE

