

Certificate of performance



Article: 3 122 107

Pellet-burettes, 10 ml, lat. needle valve, AS, conf. cert.




Batch: 02-24

Serial:

Measuring Nr. 66227

Batch size: 1

Volume:	10,000 ml	6,000 ml	2,000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Measured value:	9,9943 ml	6,0076 ml	1,9999 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	05.03.2024
Balance:	Sartorius - LE244S-OCE

