

Certificate of performance



Article: 3 122 258

Pellet-burettes, 25 ml, lat. needle valve, AS, conf. cert.




Batch: 11-22

Serial:

Measuring Nr.: 62505

Batch size: 20

Volume:	25,0000 ml	15,0000 ml	5,0000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	24,9945 ml	15,0026 ml	5,0134 ml
Deviation:	0,0097 ml	0,0115 ml	0,0008 ml
Precision (CV):	0,0388 %	0,0764 %	0,0170 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	06.04.2023
Balance:	Sartorius - LE244S-OCE

