

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 122 504

Pellet-burettes, 50 ml, lateral stopc., AS, conf. cert.




Batch: 03-17

Serial:

Measuring Nr. 17128

Batch size: 2

Volume:	50,000 ml	30,000 ml	10,000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	50,0162 ml	30,0049 ml	10,0218 ml
Deviation:	0,0119 ml	0,0150 ml	0,0104 ml
Precision (CV):	0,0238 %	0,0501 %	0,1036 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	17.01.2018
Balance:	Sartorius - LE244S-OCE

