

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 122 508

Pellet-burettes, 50 ml, lat. needle valve, AS, conf. cert.




Batch: 30-17

Serial:

Measuring Nr. 19678

Batch size: 20

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	49,9876 ml	30,0221 ml	10,0171 ml
Deviation:	0,0138 ml	0,0131 ml	0,0112 ml
Precision (CV):	0,0276 %	0,0437 %	0,1118 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	10.04.2018
Balance:	Sartorius - LE244S-OCE

DE-M