

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 122 508

Pellet-burettes, 50 ml, lat. needle valve, AS, conf. cert.




Batch: 17-24

Serial:

Measuring Nr. 77298

Batch size: 10

Volume:	50,000 ml	30,000 ml	10,000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	50,0087 ml	29,9963 ml	10,0145 ml
Deviation:	0,0133 ml	0,0082 ml	0,0086 ml
Precision (CV):	0,0267 %	0,0273 %	0,0860 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	08.04.2026
Balance:	Sartorius - LE244S-OCE

