

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 124 105

Pellet-burettes, 10 ml, amber glass, AS, conf. cert.

Batch: 29-20




Serial:

Measuring Nr. 44344

Batch size:

7

Volume:	10,000 ml	6,000 ml	2,000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	10,0154 ml	6,0050 ml	2,0016 ml
Deviation:	0,0018 ml	0,0047 ml	0,0050 ml
Precision (CV):	0,0180 %	0,0787 %	0,2487 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

77

Date:

09.02.2021

Balance:

Sartorius -
LE244S-OCE

