

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 124 106

Pellet-burettes, 10 ml, amber glass, AS, conf. cert.




Batch: 47-15

Serial:

Measuring Nr. 13951

Batch size: 1

Volume:	10,000 ml	6,000 ml	2,000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Measured value:	10,0101 ml	6,0090 ml	1,9996 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	07.09.2017
Balance:	Sartorius - LE244S-OCE

