

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 124 108

Pellet-burettes, 10 ml, amber glass, AS, conf. cert.




Batch: 47-15

Serial:

Measuring Nr.: 16432

Batch size: 10

Volume:	10,000 ml	6,000 ml	2,000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	9,9900 ml	6,0015 ml	2,0070 ml
Deviation:	0,0012 ml	0,0058 ml	0,0011 ml
Precision (CV):	0,0119 %	0,0974 %	0,0543 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	11.12.2017
Balance:	Sartorius - LE244S-OCE

