

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 124 505

Pellet-burettes, 50 ml, amber glass, AS, conf. cert.




Batch: 18-21

Serial:

Measuring Nr. 52147

Batch size: 21

Volume:	50,0000 ml	30,0000 ml	10,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	50,0139 ml	30,0144 ml	10,0088 ml
Deviation:	0,0096 ml	0,0150 ml	0,0117 ml
Precision (CV):	0,0192 %	0,0500 %	0,1169 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	16.02.2022
Balance:	Sartorius - LE244S-OCE

