

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 124 505

**Pellet-burettes, 50 ml, amber glass, AS, conf. cert.**




**Batch:** 18-24

**Serial:**

**Measuring Nr.** 69126

**Batch size:** 33

Volume:	50,000 ml	30,000 ml	10,000 ml
<b>Tolerance:</b>	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
<b>Average value:</b>	49,9969 ml	30,0257 ml	10,0289 ml
<b>Deviation:</b>	0,0124 ml	0,0088 ml	0,0035 ml
<b>Precision (CV):</b>	0,0248 %	0,0293 %	0,0348 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	16.09.2024
<b>Balance:</b>	Sartorius - LE244S-OCE

