Certificate of performance


Article:	**3 154 255**
Universal autom. burettes, 25 ml, AS, conf. cert.	


witeg GERMANY


ISO 9001 CERTIFIED

Batch:	48-22	**Serial:**	
Measuring Nr.	80384	**Batch size:**	1

Volume:	**25,0000 ml**	**15,0000 ml**	**5,0000 ml**
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Measured value:	25,0194 ml	14,9967 ml	5,0008 ml







These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH
Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	28.10.2026
Balance:	Sartorius - LE244S-0CE

DE-M