

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 372 053

Micro burette, 5 ml, str. stopc., PTFE, AS, conf. cert.

Batch: 07-26




Serial:

Measuring Nr. 78064

Batch size:

10

Volume:	5,0000 ml	3,0000 ml	1,0000 ml
Tolerance:	$\pm 0,0100$ ml	$\pm 0,0100$ ml	$\pm 0,0100$ ml
Average value:	4,9996 ml	2,9993 ml	1,0013 ml
Deviation:	0,0027 ml	0,0016 ml	0,0029 ml
Precision (CV):	0,0538 %	0,0544 %	0,2865 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

19.05.2026

Balance:

Sartorius -
LE244S-OCE

DE-M