

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 372 102

**Micro burette, 10 ml, str. Stopc., glass, AS, conf. cert.**




**Batch:** 11-22

**Serial:**

**Measuring Nr.** 66830

**Batch size:** 10

Volume:	10,000 ml	6,000 ml	2,000 ml
Tolerance:	± 0,0200 ml	± 0,0200 ml	± 0,0200 ml
Average value:	10,0073 ml	6,0035 ml	2,0026 ml
Deviation:	0,0008 ml	0,0047 ml	0,0054 ml
Precision (CV):	0,0078 %	0,0778 %	0,2682 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

Operator:	7
Date:	17.04.2024
Balance:	Sartorius - LE244S-OCE

