

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 383 025

Micro burettes, 2 ml, lat. needle valve, AS, conf. cert.




Batch: 40-25

Serial:

Measuring Nr.: 75196

Batch size: 10

Volume:	2,0000 ml	1,2000 ml	0,4000 ml
Tolerance:	± 0,0100 ml	± 0,0100 ml	± 0,0100 ml
Average value:	1,9979 ml	1,1936 ml	0,3984 ml
Deviation:	0,0012 ml	0,0012 ml	0,0022 ml
Precision (CV):	0,0579 %	0,0991 %	0,5536 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	04.11.2025
Balance:	Sartorius - LE244S-OCE

