

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 715 001 BL

Grad. cylinders, 10:0,2 ml, blue grad, A, conf. cert.




Batch: 23-24

Serial:

Measuring Nr. 67407

Batch size: 280

Volume:	2,0000 ml	6,0000 ml	10,0000 ml
Tolerance:	± 0,1000 ml	± 0,1000 ml	± 0,1000 ml
Average value:	2,0123 ml	5,9962 ml	9,9887 ml
Deviation:	0,0189 ml	0,0188 ml	0,0213 ml
Precision (CV):	0,9455 %	0,3134 %	0,2134 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 4788

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	10.06.2024
Balance:	Sartorius - LE244S-OCE

